

# CE DECLARATION OF CONFORMITY

## CEE MEDICAL DEVICES DIRECTIVE 93/42/EEC

**PRODUCT MANUFACTURER:** BRIX S. R. L.  
**ADDRESS:** Parque Industrial Carcarañá  
Ruta 9 km 348,5  
(2138) Carcarañá – Provincia de Santa Fe - Argentina

**EUROPEAN REPRESENTATIVE:** CONCEPTOS FUNCIONES Y ESTETICA DENTAL IBERICA S.L.  
**ADDRESS:** Gran Vía 8-10 1º 5b (08902) Hospitalet de Llobregat, Barcelona, España.

**DECLARE UNDER THEIR RESPONSIBILITY THAT THE PRODUCT:**

**Name** Brix 3000

**ECRI-UMDNS Code:** 15-584 GEL

**Type:** Gel for atraumatic removal in carious lesions. Syringes of 0.5, 1.0, 3.0 or 5.0 ml or tubes of 2, 3, 4, 5 or 6 ml multidose in individual boxes. Boxes containing 25, 50 or 100 units.

**CONFORMS WITH THE REQUISITES OF THE DIRECTIVES**

**EC Directive 93/42/CEE** Medical Devices Directive 93/42/CEE amended by Directive 2007/47/CE.

**Classification (rule):** Class I (rule 5)

**Compliance with harmonized standards**

ISO 13485:2013 / UNE-EN ISO 14971:2012 / UNE-EN ISO 10993  
ANMAT Disp.Nº2318/02 T.O. 2004, ANMAT Disp.Nº 3266/13.

**OTHER INFORMATION:** A.F.E.Disp.6227/14 Legajo 2177 PM-2177-1  
Design, Development, Manufacture and Sales of gel for a traumatic removal in carious lesions of dental use Certificate No.: 245221-2017-AQ-ARG-NA-PS Project No.: PRJC-551635-2016-MS-ARG Initial certification date: 20 August 2017 Valid until: 28 February 2019

**DATE:** April 30<sup>th</sup>, 2018

	Farm. María Laura Borga	Mauricio Dobboletta
<b>Signed</b>	 MARIA LAURA BORGA FARMACEUTICA Mat. 2106	 MAURICIO DOBBOLETTA SOCIO GERENTE BRIX S.R.L.
<b>Function</b>	Safety Officer	General Manager